


TruSeq Amplicon - Cancer Panel

Protocol Guide

For Research Use Only. Not for use in diagnostic procedures.

Hybridize Oligo Pool	3
Remove Unbound Oligos	4
Extend and Ligate Bound Oligos	5
Amplify Libraries	6
Clean Up Libraries	7
Normalize Libraries	8
Pool Libraries	9
Acronyms	10
Technical Assistance	11



ILLUMINA PROPRIETARY

illumina®

This document and its contents are proprietary to Illumina, Inc. and its affiliates ("Illumina"), and are intended solely for the contractual use of its customer in connection with the use of the product(s) described herein and for no other purpose. This document and its contents shall not be used or distributed for any other purpose and/or otherwise communicated, disclosed, or reproduced in any way whatsoever without the prior written consent of Illumina. Illumina does not convey any license under its patent, trademark, copyright, or common-law rights nor similar rights of any third parties by this document.

The instructions in this document must be strictly and explicitly followed by qualified and properly trained personnel in order to ensure the proper and safe use of the product(s) described herein. All of the contents of this document must be fully read and understood prior to using such product(s).

FAILURE TO COMPLETELY READ AND EXPLICITLY FOLLOW ALL OF THE INSTRUCTIONS CONTAINED HEREIN MAY RESULT IN DAMAGE TO THE PRODUCT(S), INJURY TO PERSONS, INCLUDING TO USERS OR OTHERS, AND DAMAGE TO OTHER PROPERTY.

ILLUMINA DOES NOT ASSUME ANY LIABILITY ARISING OUT OF THE IMPROPER USE OF THE PRODUCT(S) DESCRIBED HEREIN (INCLUDING PARTS THEREOF OR SOFTWARE).

© 2016 Illumina, Inc. All rights reserved.

Illumina, 24sure, BaseSpace, BeadArray, BlueFish, BlueFuse, BlueGnome, cBot, CSPro, CytoChip, DesignStudio, Epicentre, ForenSeq, Genetic Energy, GenomeStudio, GoldenGate, HiScan, HiSeq, HiSeq X, Infinium, iScan, iSelect, MiniSeq, MiSeq, MiSeqDx, MiSeq FGx, NeoPrep, NextBio, Nextera, NextSeq, Powered by Illumina, SureMDA, TruGenome, TruSeq, TruSight, Understand Your Genome, UYG, VeraCode, verifi, VeriSeq, the pumpkin orange color, and the streaming bases design are trademarks of Illumina, Inc. and/or its affiliate(s) in the U.S. and/or other countries. All other names, logos, and other trademarks are the property of their respective owners.

Patent pending for methods performed by components in this kit.

For Research Use Only– not for any clinical or therapeutic use in humans or animals.
This product includes GoTaq® Hot Start Polymerase manufactured by Promega Corporation for distribution by Illumina, Inc. Licensed to Promega Corporation under U.S. Patent Nos. 5,338,671 and 5,587,287 and their corresponding foreign patents.



Hybridize Oligo Pool



WARNING

This set of reagents contains formamide, an aliphatic amide that is a probable reproductive toxin. Personal injury can occur through inhalation, ingestion, skin contact, and eye contact. Wear protective equipment, including eye protection, gloves, and laboratory coat. Handle used reagents as chemical waste and discard in accordance with the governmental safety standards for your region. For environmental, health, and safety information, see the SDS for this kit at support.illumina.com/sds.html.

Preparation

- 1 Set a 96-well heat block to 95°C.
- 2 Preheat an incubator to 37°C to prepare for the extension-ligation step.

Procedure

- 1 Add 5 µl ACD1 to 1 well of the HYP plate.
- 2 Add 5 µl gDNA to each remaining well.
- 3 Add 5 µl ACP1 to the well containing ACD1.
- 4 Add 5 µl AFP1 to each well containing gDNA.
- 5 Centrifuge at 1000 × g for 1 minute.
- 6 Add 40 µl OHS1. Pipette to mix.
- 7 Centrifuge at 1000 × g for 1 minute.
- 8 Place on the preheated heat block and incubate for 1 minute.
- 9 With the plate on the heat block, reset the temperature to 40°C and continue incubating for 80 minutes.

Remove Unbound Oligos



WARNING

This set of reagents contains formamide, an aliphatic amide that is a probable reproductive toxin. Personal injury can occur through inhalation, ingestion, skin contact, and eye contact. Wear protective equipment, including eye protection, gloves, and laboratory coat. Handle used reagents as chemical waste and discard in accordance with the governmental safety standards for your region. For environmental, health, and safety information, see the SDS for this kit at support.illumina.com/sds.html.

Preparation

- 1 Assemble the filter plate unit (FPU) from top to bottom as follows: lid, filter plate, adapter collar, and midi plate.
- 2 Wash the wells to be used in the assay with 45 μ l SW1.

Procedure

- 1 Make sure that the heat block has cooled to 40°C.
- 2 Remove from the heat block.
- 3 Centrifuge at 1000 \times g for 1 minute.
- 4 Transfer each sample to the FPU plate.
- 5 Cover and centrifuge at 2400 \times g for 2 minutes.
- 6 Wash 2 times with 45 μ l SW1.
- 7 Discard flow-through.
- 8 Reassemble the FPU plate.
- 9 Add 45 μ l UB1.
- 10 Cover and centrifuge at 2400 \times g for 2 minutes.

Extend and Ligate Bound Oligos

Procedure

- 1 Add 45 μ l ELM3 to the FPU plate.
- 2 Incubate at 37°C for 45 minutes.

Amplify Libraries

Preparation

- 1 Save the following PCR program on a thermal cycler.
 - ▶ 95°C for 3 minutes
 - ▶ 27 cycles of:
 - ▶ 95°C for 30 seconds
 - ▶ 62°C for 30 seconds
 - ▶ 72°C for 60 seconds
 - ▶ 72°C for 5 minutes
 - ▶ Hold at 10°C

Procedure

- 1 Arrange the Index 1 (i7) adapters in columns 1–12 of the TruSeq Index Plate Fixture.
- 2 Arrange the Index 2 (i5) adapters in rows A–H of the TruSeq Index Plate Fixture.
- 3 Place the plate on a TruSeq Index Plate Fixture.
- 4 Using a multichannel pipette, add 4 μ l of each Index 1 (i7) adapter down each column.
- 5 Using a multichannel pipette, add 4 μ l of each Index 2 (i5) adapter across each row.
- 6 Remove the FPU plate from the incubator and do the following.
 - a Replace the aluminum foil seal with the filter plate lid.
 - b Centrifuge at $2400 \times g$ for 2 minutes.
 - c Add 25 μ l 50 mM NaOH. Pipette to mix.
 - d Incubate at room temperature for 5 minutes.
- 7 Add 56 μ l TDP1 to a full tube (2.8 ml) of PMM2. Invert to mix.
- 8 Transfer 22 μ l PMM2/TDP1 mixture to the IAP plate.
- 9 Transfer eluted samples from the FPU plate to the IAP plate.
- 10 Centrifuge at $1000 \times g$ for 1 minute.
- 11 Transfer the IAP plate to the post-amplification area.
- 12 Place on the preprogrammed thermal cycler and run the PCR program.

SAFE STOPPING POINT

If you are stopping, seal the plate and store at 2°C to 8°C for up to 2 days. Alternatively, leave on the thermal cycler overnight.

Clean Up Libraries

Procedure

- 1 Centrifuge the IAP plate at $1000 \times g$ for 1 minute.
- 2 Run an aliquot of library and control on 4% agarose gel (5 μ l) or Bioanalyzer (1 μ l).
- 3 Add 45 μ l AMPure XP beads to the CLP plate.
- 4 Transfer all the supernatant from the IAP plate to the CLP plate.
- 5 Shake at 1800 rpm for 2 minutes.
- 6 Incubate at room temperature for 10 minutes.
- 7 Place on a magnetic stand and wait until the liquid is clear (~2 minutes).
- 8 Remove and discard all supernatant from each well.
- 9 Wash 2 times with 200 μ l 80% EtOH.
- 10 Use a 20 μ l pipette to remove residual EtOH.
- 11 Remove from the magnetic stand and air-dry for 10 minutes.
- 12 Add 30 μ l EBT.
- 13 Shake at 1800 rpm for 2 minutes.
- 14 Incubate at room temperature for 2 minutes.
- 15 Place on a magnetic stand and wait until the liquid is clear (~2 minutes).
- 16 Transfer 20 μ l supernatant from the CLP plate to the LNP plate.
- 17 Centrifuge at $1000 \times g$ for 1 minute.

Normalize Libraries



WARNING

This set of reagents contains formamide, an aliphatic amide that is a probable reproductive toxin. Personal injury can occur through inhalation, ingestion, skin contact, and eye contact. Wear protective equipment, including eye protection, gloves, and laboratory coat. Handle used reagents as chemical waste and discard in accordance with the governmental safety standards for your region. For environmental, health, and safety information, see the SDS for this kit at support.illumina.com/sds.html.



WARNING

This set of reagents contains β -mercaptoethanol. Perform the following procedure in a hood or well-ventilated area.

Procedure

- 1 For 96 samples, add 4.4 ml LNA1 to a new 15 ml conical tube.
- 2 Use a P1000 pipette to resuspend LNB1.
- 3 Transfer 800 μ l LNB1 to the tube of LNA1.
- 4 Add 45 μ l LNA1/LNB1 to the LNP plate.
- 5 Shake at 1800 rpm for 30 minutes.
- 6 Place on a magnetic stand and wait until the liquid is clear (~2 minutes).
- 7 Remove and discard all supernatant.
- 8 Remove from the magnetic stand.
- 9 Wash 2 times with 45 μ l LNW1.
- 10 Use a 20 μ l pipette to remove residual LNW1.
- 11 Remove from the magnetic stand.
- 12 Add 30 μ l fresh 0.1 N NaOH.
- 13 Shake at 1800 rpm for 5 minutes.
- 14 Place the LNP plate on a magnetic stand and wait until the liquid is clear (~2 minutes).
- 15 Add 30 μ l LNS1 to the SGP plate.
- 16 Transfer 30 μ l supernatant from the LNP plate to the SGP plate.
- 17 Centrifuge at 1000 \times g for 1 minute.

SAFE STOPPING POINT

If you are stopping, seal the plate and store at -25°C to -15°C for up to 30 days.

Pool Libraries

Procedure

- 1 Centrifuge at $1000 \times g$ for 1 minute.
- 2 Transfer 5 μ l of each library to an 8-tube strip, column by column.
- 3 Transfer the contents of the 8-tube strip to the PAL tube. Pipette to mix.
- 4 Denature and dilute pooled libraries to the loading concentration for the sequencing instrument you are using. See the denature and dilute libraries guide for your instrument.

Acronyms

Acronym	Definition
ACD1	Amplicon Control DNA 1
ACP1	Amplicon Control Oligo Pool 1
AFP1	Amplicon Fixed Panel 1
CLP	Clean-up Plate
EBT	Elution Buffer with Tris
ELM3	Extension Ligation Mix 3
FPU	Filter Plate Unit
HT1	Hybridization Buffer
HYP	Hybridization Plate
IAP	Indexed Amplification Plate
LNA1	Library Normalization Additives 1
LNB1	Library Normalization Beads 1
LNS1	Library Normalization Storage Buffer 1
LNW1	Library Normalization Wash 1
LNP	Library Normalization Plate
OHS1	Oligo Hybridization for Sequencing Reagent 1
PAL	Pooled Amplicon Library
PMM2	PCR Master Mix 2
SGP	Storage Plate
SW1	Stringent Wash 1
TDP1	TruSeq DNA Polymerase 1
UB1	Universal Buffer 1

Technical Assistance

For technical assistance, contact Illumina Technical Support.

Table 1 Illumina General Contact Information

Website	www.illumina.com
Email	techsupport@illumina.com

Table 2 Illumina Customer Support Telephone Numbers

Region	Contact Number	Region	Contact Number
North America	1.800.809.4566	Japan	0800.111.5011
Australia	1.800.775.688	Netherlands	0800.0223859
Austria	0800.296575	New Zealand	0800.451.650
Belgium	0800.81102	Norway	800.16836
China	400.635.9898	Singapore	1.800.579.2745
Denmark	80882346	Spain	900.812168
Finland	0800.918363	Sweden	020790181
France	0800.911850	Switzerland	0800.563118
Germany	0800.180.8994	Taiwan	00806651752
Hong Kong	800960230	United Kingdom	0800.917.0041
Ireland	1.800.812949	Other countries	+44.1799.534000
Italy	800.874909		

Safety data sheets (SDSs)—Available on the Illumina website at support.illumina.com/sds.html.

Product documentation—Available for download in PDF from the Illumina website. Go to support.illumina.com, select a product, then select **Documentation & Literature**.



Illumina

5200 Illumina Way

San Diego, California 92122 U.S.A.

+1.800.809.ILMN (4566)

+1.858.202.4566 (outside North America)

techsupport@illumina.com

www.illumina.com