

# Certificate



Certificate no.

CU 72180546 01

**License Holder:**

Illumina, Inc.  
5200 Illumina Way  
San Diego CA 92122  
USA

**Manufacturing Plant:**

Illumina Singapore Pte. Ltd.  
North Tech Lobby 3 #02-13/18  
29 Woodlands Industrial Park E1  
Singapore 757716  
Singapore

Test report no.: USA-AS 31593844 010

Client Reference: Carol Escano

Tested to:

UL 61010-1:2012  
UL 61010-2-081:2015  
CAN/CSA-C22.2 NO. 61010-1-12  
CAN/CSA-C22.2 NO. 61010-2-081:15

Certified Product: Genetic Sequencer

License Fee - Units

Model Designation: MiniSeq

7


Rated Voltage: AC 100-240V, 50/60Hz  
Rated Power: 240W  
Protection Class: I

Special Remarks: Replaces Certificate CU72153075.

7

Appendix: 1, 1-4

Licensed Test mark:



Date of Issue

(day/mo/yr)

21/03/2018